



BRUKER

Current Data Parameters
NAME sdns
EXPNO 41
PROCNO 1

F2 - Acquisition Parameters
Date_ 20181109
Time 10.35
INSTRUM spect
PROBHD 5 mm PABBO BB-
PULPROG zg30
TD 32768
SOLVENT DMSO
NS 16
DS 0
SWH 9980.040 Hz
FIDRES 0.504567 Hz
AQ 1.6417269 sec
RG 64
DW 50.100 usec
DE 6.00 usec
TE 296.5 K
D1 2.00000000 sec
TD0 1

==== CHANNEL f1 ====
NUC1 1H
PI 7.60 usec
PL1 -3.00 dB
SFO1 400.2030815 MHz

F2 - Processing parameters
SI 16384
SF 400.2000348 MHz
WDW EM
SGB 0
LB 0.00 Hz
GB 0
PC 1.00