



LCMS OF 1187594-09-7

```

File D:\DATA\A02\MONITOR\BRTN1-00381-LCMSA035.D
Injection Date: 25 Apr 1:18 pm +0800 Tgt Mass (E2X) :
Sample Name : BRTN Location : P1-C-08
Acq. Operator : A02-MONITOR Inj : 1
Spec. Reported : MS Integration Inj Volume : 1 ul
Acq. Method : D:\METHODS\1-POS-3MIN.M
Analysis Method : D:\METHODS\1-POS-3MIN.M
Sample Info : Easy-Access Method: '1-POS-3MIN'
Method Info : Mobile Phase: A: water(0.01%TFA) B: ACN(0.01%TFA)
Gradient: 5%-95% B in 1.5min
Flow Rate: 1.5 ml/min
Column: SunFire C18, 3*30mm, 2.5um
Oven Temperature: 50°C
Mass Range: 110-1500
    
```



Integration Results for ADC1 A, ELSD


RetTim	Width	Area	Height	Area%
0.75	0.02	64.26	42.39	100.00

Integration Results for DAD1 A, Sig=214.4 Ref:off

RetTim	Width	Area	Height	Area%
0.75	0.02	3338.88	2456.07	100.00

Integration Results for DAD1 B, Sig=254.4 Ref:off

RetTim	Width	Area	Height	Area%
0.75	0.02	2342.59	1850.68	100.00



* If you have any comments and suggestions, please contact us by sales@amadischem.com