


HPLC of 875337-44-3

HPLC REPORT			
Sample Name:	MG265	Injection Volume:	10.0
Vial Number:	RE4	Channel:	UV_VIS_3
Sample Type:	unknown	Wavelength:	254
Control Program:		Bandwidth:	4
Quantif. Method:	default	Dilution Factor:	1.0000
Recording Time:	2015/12/5 13:07	Sample Weight:	1.0000
Run Time (min):	13.19	Sample Amount:	1.0000



No.	Ret. Time min	Peak Name	Height mAU	Area mAU*min	Rel. Area %	Amount n.a.	Type
1	1.74	n.a.	7.432	0.809	0.32	n.a.	BMB*
2	1.83	n.a.	0.131	0.011	0.00	n.a.	Rd
3	2.62	n.a.	3.036	0.366	0.14	n.a.	BM*
4	2.89	n.a.	2.539	0.436	0.17	n.a.	M*
5	3.26	n.a.	3.740	0.620	0.24	n.a.	M*
6	3.76	n.a.	6.708	1.426	0.56	n.a.	MB*
7	4.59	n.a.	1116.973	250.381	98.56	n.a.	BMB*
Total:			1140.561	254.050	100.00	0.000	

* If you have any comments and suggestions, please contact us by sales@amadischem.com