


### HPLC of 147732-56-7

Column : YMC-Pack ODS-AQ, 4.6\*250mm, 5um.  
 Solvent A : 0.1%Trifluoroacetic in 100% Acetonitrile  
 Solvent B : 0.1%Trifluoroacetic in 100% Water  
 Gradient :  
 0.0min 47% 53%  
 20.0min 67% 33%  
 20.1min 100% 0%  
 25.0min Stop  
 Flow rate : 1.0ml/min  
 Wavelength : 220nm



Rank	Time	Name	Conc.	Area
1	8.575		0.2227	58225
2	8.845		98.96	25877358
3	9.833		0.6853	179196
4	11.199		0.1338	34988
Total			100	26149764

\*if you have any comments and suggestions, please contact us by [sales@amadischem.com](mailto:sales@amadischem.com)